

Certificate of Quality

Product Name: Nuclease AWAY

Product Code: A6257 CAS #:
Grade: Biotechnology Formula:
Appearance: MW:
Reassay Date: 3 years after shipping date Lot#: See product label
Storage: Store at room temperature (+15°C).

Laboratory glassware was intentionally contaminated with Nuclease, then treated with Nuclease away. After which the glassware was subjected to contamination of deionized water rinses and wiping dry. The pieces were extracted with DEPC water, and the extracts were then added to an RNA standard and incubated. Following incubation, the samples were analyzed by gel electrophoresis and evaluated for degradation of the RNA. Test results confirm that, when used as directed this lot of RNase away is effective at removing nuclease contamination from laboratory surfaces.

TEST	SPECIFICATION	RESULT
Appearance	Clear solution	Passes

Data Verified by M. He. Verification Date: 2020-01-06

Limited Usage and Warranty: For Research Use Only. Not for use in diagnostic procedures. Not for human or animal drug or food use. Biomatik warrants materials of said quality at the time of sale. It is the sole responsibility of the customer to determine the adequacy of all materials for any intended or specific purpose of use. Biomatik's sole obligation is to replace the material up to the extent of the purchase price. This warranty applies only to products in original packaging and does not apply to a product which has been tampered with or altered in any way or which has been misused or damaged by accident or negligence. No other warranties of any kind, expressed or implied, including without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Biomatik. Biomatik shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.