



RPU42664 Human 200µg
Recombinant Major Histocompatibility Complex Class I E (MHCE)
Organism Species: *Homo sapiens (Human)*
Instruction manual

FOR RESEARCH USE ONLY
NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

12th Edition (Revised in Aug, 2016)



[PROPERTIES]

Source: Prokaryotic expression

Host: *E.coli*

Residues: Gly22~Leu358

Tags: N-terminal His Tag

Subcellular Location: Membrane

Purity: > 95%

Traits: Freeze-dried powder

Buffer formulation: 20mM Tris, 150mM NaCl, pH8.0, containing 1mM EDTA, 0.01% SKL, 5% Trehalose and Proclin300.

Original Concentration: 200µg/mL

Applications: Positive Control; Immunogen; SDS-PAGE; WB.

(May be suitable for use in other assays to be determined by the end user.)

Predicted isoelectric point: 5.8

Predicted Molecular Mass: 41.5kDa

Accurate Molecular Mass: 42kDa as determined by SDS-PAGE reducing conditions.

[USAGE]

Reconstitute in 20mM Tris, 150mM NaCl (pH8.0) to a concentration of 0.1-1.0 mg/mL. Do not vortex.

[STORAGE AND STABILITY]

Storage: Avoid repeated freeze/thaw cycles.

Store at 2-8°C for one month.

Aliquot and store at -80°C for 12 months.

Stability Test: The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.



[SEQUENCE]

```
          GSHSLKYFH TSVSRPGRGE PRFISVGYVD  
DTQFVRFDND AASPRMVPRA PWMEQEGSEY WDRETRSARD TAQIFRVNLR  
TLRGYYNQSE AGSHTLQWMH GCELGPDRRF LRGYEQFAYD GKDYLTNED  
LRSWTAVDTA AQISEQKSND ASEAEHQRAY LEDTCVEWLH KYLEKGKETL  
LHLEPPKTHV THHPISDHEA TLRCWALGFY PAEITLTWQQ DGEGHTQDTE  
LVETRPAGDG TFQKWAAVVV PSGEEQRYTC HVQHEGLPEP VTLRWKPASQ  
PTIPIVGIIA GLVLLGSVVS GAVVAAVIWR KKSSGGKGGG YSKAEWSDSA  
QGSESHSL
```

[IDENTIFICATION]




Figure. Gene Sequencing (Extract)




Figure. SDS-PAGE

[IMPORTANT NOTE]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.